

UNDERWRITERS LABORATORIES INC.®

CERTIFICATE OF COMPLIANCE



(1)

(2) UL Certificate Number 03.21968

(3) This certificate is issued for: Toxic Gas Detectors
Type/Model: GasAlertClip Extreme Series
Models GA24A-H-DL-XX, GA24A-M-DL-XX,
GA24A-S-DL-XX and GA24A-X-DL-XX

(4) Issued to: BW Technologies Ltd
2840 2nd Avenue S.E.
Calgary AB, CANADA T2A 7X9

(5) Report Reference: E131021, Issued August 28, 2003

(6) This product has been investigated by Underwriters Laboratories Inc.® in accordance with the following Standard(s) for Safety indicated on this Certificate:

UL 913
CSA C22.2 No. 157-92

(7) The product marking shall include the following information:

Class I, Groups A, B, C, and D, T4
Class I, Zone 0, Group IIC T4

(8) Only those products bearing the UL Classification Mark for the US and Canada should be considered as being covered by UL's Classification and Follow-Up Service meeting the appropriate requirements for the US and Canada.

The UL Classification Mark for the US and Canada generally includes: the UL in a circle symbol with "C" and "US" identifiers; with the word "CLASSIFIED" (as shown); a control number (may be alphanumeric) assigned by UL; a statement to indicate the extent of UL's evaluation of the product; and the product category name (product identifier) as indicated in the appropriate UL Directory.

LOOK FOR THE UL CLASSIFICATION MARK ON THE PRODUCT

Signed:

Paul Kelly
Manager

Issued: 1 October 2003

This certificate may only be reproduced in its entirety and without alterations.